

The Performance of a Continuous Supersonic Expansion Discharge Source

Carrie A. Kauffman (kauffma2@illinois.edu), Kyle N. Crabtree, *Department of Chemistry, University of Illinois, Urbana, IL 61801*; Benjamin J. McCall, *Departments of Chemistry and Astronomy, University of Illinois, Urbana, IL 61801*.

Production of cold molecular ions is an integral component in performing high resolution spectroscopy of astrophysically important ions at temperatures comparable to the interstellar medium. Typically, pulsed supersonic expansion discharge sources have been used to produce ions at these temperatures. However, pulsed sources impose a limit on the duty cycle of the spectroscopic technique. To circumvent this limitation, a continuous supersonic expansion discharge source has been designed for use in the high resolution technique that is currently being developed in our lab, SCRIBES (Sensitive, Cooled, Resolved Ion BEam Spectroscopy). In order to characterize and optimize the source design, the $R(1,0)$, $R(1,1)^u$, and $R(2,2)^l$ transitions within the ($\nu_2 \leftarrow 0$) band of H_3^+ have been studied using continuous-wave cavity ringdown spectroscopy. Initial performance results will be discussed.

